



SAFETY AND BUILDINGS DIVISION
Plumbing Product Review
P.O. Box 2658
Madison, Wisconsin 53701-2658
TTY: Contact Through Relay

Scott Walker, Governor
Dave Ross, Secretary

April 3, 2012

MONROE CLINIC
STEVE BOROWSKI
515 22ND AVE.
MONROE WI 53566

CS MEDICAL, LLC
LEE GILIAM
2150 E. EAST WILTON AVE.
CREEDMORE NC 27522

Re: Description: HEALTH CARE PLUMBING APPLIANCE
Manufacturer: CS MEDICAL, LLC
Product Name: TEE PROBE DISINFECTOR
Model Number(s): TD-100
Product File No: 20120109

The specifications and/or plans for this plumbing product have been reviewed and determined to be in compliance with chapters SPS 382 through 384, Wisconsin Administrative Code, and Chapters 145 and 160, Wisconsin Statutes.

The Department hereby issues an approval based on the Wisconsin Statutes and the Wisconsin Administrative Code. This approval is valid until the end of April 2017.

This approval is contingent upon compliance with the following stipulation(s):

- This product shall have a high hazard assembly installed on the water line. (RPZ ASSE 1013, SVB ASSE 1056)

The drain on this product shall be installed in accordance to the manufactures recommendations.

The department is in no way endorsing this product or any advertising, and is not responsible for any situation which may result from its use.

Sincerely,

Jerry Thompson
Plumbing Product Reviewer
Phone: 608-266-6742
Fax: 608-267-9566
E-mail: jerry.thompson@wi.gov